

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAR 02 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130493
Name	CHRISTOPHER WILLIAMS	ID#	7096
Date	02/22/2010		
A	Agency	Clear Security Forces	Phone # (907) 585-6293
	Instrument Location	Bldg 37, Clear AFS	
B	Alco S/N	X124609	Target Value .081
			High Pressure 1175
	Alco Test Values	.075	.077
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature	<i>Ch. Williams</i>	<i>CAB</i> 3/4/10
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of March, 2010.

*Carolyn M. Noland*  
(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

BT9 06/03

